

临床生化质控血清出厂检验报告

检验依据: 临床生化质控血清说明书 检验标准: 20172400256

项目	检测值	批内 CV	批间 CV	判定
要求	(详见靶值单)。	CV≤5%	CV≤8%	合格
ALT/GPT	65	1.65%	1.75%	√
GOT/AST	70	1.67%	0.33%	√

SP320000	130	2.37%	2.47%	√
LDH	138	0.88%	1.69%	√
CK	133	3.33%	1.39%	√
α-AMY	138	0.82%	0.45%	√
TP	33.3	0.44%	0.34%	√
ALP	38.3	1.87%	0.46%	√
Crea (肌酐酶法)	122	3.83%	1.88%	√
UA (尿酸酶法)	431	3.84%	2.37%	√
UA	292	1.83%	0.25%	√
Urea	6.20	1.84%	1.53%	√
TG	1.11	2.88%	1.87%	√
CHO	3.98	4.53%	3.54%	√
TBA	29.5	1.27%	0.33%	√
TB (重氮法)	24.3	1.37%	0.33%	√
TB (钼酸盐法)	30.3	0.87%	0.43%	√
DB (孔雀绿)	13.3	1.88%	0.38%	√
DB (钼酸盐法)	18.3	2.11%	0.48%	√
Ch (胆固醇法)	1.67	3.87%	2.17%	√
F	1.38	2.58%	3.77%	√
Cl (汞电极法)	81	1.18%	0.48%	√
K	3.80	1.18%	0.21%	√
Na	116	0.85%	0.21%	√

检验人:

复核人:

